

FID: Free Induction Decay

- The Received NMR Signal is a Decaying Sinewave
- The Frequency of this Sinewave is the Peak Position
- Many Peaks in the Spectrum Means Many Sinewaves of Different Frequency Added Together in the FID
- The Fourier Transform Extracts the Frequency Information from the FID