

Digitized FID Data, (DW=80)

(μ s) <u>Time</u>	Varian (simultaneous)		Bruker (alternate)	
0	1. 2345	2. -227	1. 23465	
80				2. 9355
160	3. 142	4. 3481	3. -746	
240				4. -1793
320	5. -263	6. -8951	5. 472	
400				6. 6758
	<u>real</u>	<u>imag</u>	<u>real</u>	<u>imag</u>

FID files:

Varian: 2345, -227, 142, 3481, -263, -8951

Bruker: 23465, 9355, -746, -1793, 472, 6758
real, imag, real, imag, real, imag